

106530707

PCT/DE2003/003164



Rec'd 10 APR 2003

PATENT COOPERATION TREATY

PCT

Translation

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY
(Chapter II of the Patent Cooperation Treaty)

(PCT Article 36 and Rule 70)

Applicant's or agent's file reference K-258 PCT	FOR FURTHER ACTION See Form PCT/IPEA/416	
International application No. PCT/DE2003/003164	International filing date (day/month/year) 24 September 2003 (24.09.2003)	Priority date (day/month/year) 10 October 2002 (10.10.2002)
International Patent Classification (IPC) or national classification and IPC G01B 17/02		
Applicant AGFA NDT GMBH		

1. This report is the international preliminary examination report, established by this International Preliminary Examining Authority under Article 35 and transmitted to the applicant according to Article 36.

2. This REPORT consists of a total of 4 sheets, including this cover sheet.

3. This report is also accompanied by ANNEXES, comprising:

a. ☐ (sent to the applicant and to the International Bureau) a total of _____ sheets, as follows:

☐ sheets of the description, claims and/or drawings which have been amended and are the basis of this report and/or sheets containing rectifications authorized by this Authority (see Rule 70.16 and Section 607 of the Administrative Instructions).

☐ sheets which supersede earlier sheets, but which this Authority considers contain an amendment that goes beyond the disclosure in the international application as filed, as indicated in item 4 of Box No. I and the Supplemental Box.

b. ☐ (sent to the International Bureau only) a total of (indicate type and number of electronic carrier(s)) _____, containing a sequence listing and/or tables related thereto, in computer readable form only, as indicated in the Supplemental Box Relating to Sequence Listing (see Section 802 of the Administrative Instructions).

4. This report contains indications relating to the following items:

☒ Box No. I Basis of the report

☐ Box No. II Priority

☐ Box No. III Non-establishment of opinion with regard to novelty, inventive step and industrial applicability

☐ Box No. IV Lack of unity of invention

☒ Box No. V Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement

☐ Box No. VI Certain documents cited

☐ Box No. VII Certain defects in the international application

☐ Box No. VIII Certain observations on the international application

Date of submission of the demand 20 January 2004 (20.01.2004)	Date of completion of this report 21 December 2004 (21.12.2004)
Name and mailing address of the IPEA/EP	Authorized officer
Facsimile No.	Telephone No.

INTERNATIONAL PRELIMINARY REPORT ON PATENTABILITY

International application No.

PCT/DE2003/003164

Box No. I Basis of the report

1. With regard to the language, this report is based on the international application in the language in which it was filed, unless otherwise indicated under this item.

- ☐ This report is based on translations from the original language into the following language _____, which is language of a translation furnished for the purpose of:
- ☐ international search (under Rules 12.3 and 23.1(b))
 - ☐ publication of the international application (under Rule 12.4)
 - ☐ international preliminary examination (under Rules 55.2 and/or 55.3)

2. With regard to the elements of the international application, this report is based on (replacement sheets which have been furnished to the receiving Office in response to an invitation under Article 14 are referred to in this report as "originally filed" and are not annexed to this report):

- ☐ The international application as originally filed/furnished
- ☒ the description:
- pages _____ 1-14 _____, as originally filed/furnished
- pages* _____ received by this Authority on _____
- pages* _____ received by this Authority on _____
- ☒ the claims:
- pages _____ 1-8 _____, as originally filed/furnished
- pages* _____, as amended (together with any statement) under Article 19
- pages* _____ received by this Authority on _____
- pages* _____ received by this Authority on _____
- ☒ the drawings:
- pages _____ 1/2-2/2 _____, as originally filed/furnished
- pages* _____ received by this Authority on _____
- pages* _____ received by this Authority on _____
- ☐ a sequence listing and/or any related table(s) – see Supplemental Box Relating to Sequence Listing.

3. ☐ The amendments have resulted in the cancellation of:

- ☐ the description, pages _____
- ☐ the claims, Nos. _____
- ☐ the drawings, sheets/figs _____
- ☐ the sequence listing (specify): _____
- ☐ any table(s) related to sequence listing (specify): _____

4. ☐ This report has been established as if (some of) the amendments annexed to this report and listed below had not been made, since they have been considered to go beyond the disclosure as filed, as indicated in the Supplemental Box (Rule 70.2(c)).

- ☐ the description, pages _____
- ☐ the claims, Nos. _____
- ☐ the drawings, sheets/figs _____
- ☐ the sequence listing (specify): _____
- ☐ any table(s) related to sequence listing (specify): _____

* If item 4 applies, some or all of those sheets may be marked "superseded."

INTERNATIONAL PRELIMINARY EXAMINATION REPORT

International application No.

PCT/DE 03/03164

V. Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement

1. Statement

Novelty (N)	Claims	2-8	YES
	Claims	1	NO
Inventive step (IS)	Claims		YES
	Claims	1-8	NO
Industrial applicability (IA)	Claims	1-8	YES
	Claims		NO

2. Citations and explanations

Reference is made to the following document:

D1: US-A-5 717 142 (SCHAFFER MARK EVAN)

10 February 1998 (1998-02-10)

Novelty

The present application does not meet the requirements of PCT Article 33(1) since the subject matter of claim 1 is not novel within the meaning of PCT Article 33(2).

D1 discloses: an ultrasound test device with a transmitter, a receiver and a monitor. The ultrasound test device is an apparatus for determining animal backfat thickness, which is also suitable for testing a workpiece. The echo signals received by the receiver are displayed on the monitor. The same monitor also displays additional data concerning the quality of the probe-to-specimen contact (page 1, lines 20 to 42; page 4, line 54, to page 5, line 54; figures 8 and 9). Features 72 and 80 in figures 8 and 9 (**not** the display unit in the casing 4 heavily outlined in black) are covered by the term "bar-type display", i.e. a display means in the form of a bar.

Dependent claims

Dependent claims 2 to 8 do not contain any features which, combined with the features of any claim to which they refer, meet the PCT novelty and inventive step requirements. The reasons for this are that the dependent claims concern technically conventional minor structural amendments to the arrangement as per claim 1, in particular since the advantages afforded thereby can easily be foreseen. In particular, the feature in claim 5, namely the arrangement of the bar-type display, is not inventive since it does not solve a technical problem.